SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

TENDER DOCUMENT FOR

"Installation and commissioning of Mammogram Machine (Motorized Compression, Digital Display, Self-Diagnostic Interlocks) at Comprehensive Women's Preventing Health and Early Breast & Cervical Cancer Screening Center at Gautam Buddha Nagar

at

SHEWINGS FOUNDATION
UNITNO.C-403 04TH FLOOR TOWER C
TECHNOPARK PLOT NO 5, SECTOR 127,
NOIDA

NAME OF TENDERER _____ADDRESS____

Last Date of Submission;-16-10-2026

SHEWings Foundaton
UNIT NO.C-403 04TH FLOOR
TOWER CTECHNOPARK PLOT N 5,
SECTOR 127, NOIDA

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

SHEWingsFoundaton
UNIT NO.C-403 04 TH FLOOR TOWER C TECHNOPARKPLOT NO 5, SECTOR 127, NOIDA

“Installation and commissioning of Mammogram Machine (Motorized Compression, Digital Display, Self-Diagnostic Interlocks) at Comprehensive Women's Preventing Health and Early Breast & Cervical Cancer Screening Center at Gautam Buddha Nagar

at

SHEWingsFoundaton
UNIT NO.C-403 04TH FLOOR TOWER C TECHNOPARK PLOT NO 5,SECTOR 127,NOIDA

PARTI (Pre-qualificationBid)

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

NOTICE INVITING TENDER

Ref.No. SWF/MAMMO/2026

Dated:23-09-2026

Dear Sir

"Installation and commissioning of Mammogram Machine (Motorized Compression, Digital Display, Self-Diagnostic Interlocks) at Comprehensive Women's Preventing Health and Early Breast & Cervical Cancer Screening Center at Gautam Buddha Nagar

At

SHEWINGSFOUNDATION
UNIT NO.C-403 04 TH FLOOR TOWER C, TECHNOPARK PLOT NO-5, SECTOR 127, NOIDA

Tenders are invited from eligible interested parties to tender for the captioned work. The tender form can be obtained from the office of Chief Planning Officer, SHEWingsFoundaton **UNITNO.C-403 04THFLOOR TOWERCTECHNOPARKPLOT NO5, SECTOR 127, NOIDA** on or before 16-10-2026 or may be downloaded from SHEWINGSFOUNDATION website (www.SheWings.com/tenders)after satisfying the fulfillment of the criteria mentioned therein.

Please Submit tender document in the prescribed tender form in envelopes as follows should be addressed to Chief Planning Officer , SHEWINGS FOUNDATION UNIT NO.C-403 04TH FLOOR TOWER C TECHNOPARK PLOT NO 5, SECTOR 127, NOIDA,

The envelope shall be dropped in the Tender Box provided at the unit no. C-403, 4th floor Technopark sector 127 Noida. not later than 14.00 Hrs on 16-10-2026.

If the last date of receipt or opening of the tenders happens to be a holiday for SHEWINGS FOUNDATION ,then the receipt and opening by the tenders shall be shifted to next working day without change of time and venue.

SHEWINGSFOUNDATION reserves the right to accept or reject any tender either in whole, or in part without assigning any reasons for doing so and do not bound themselves to accept the lowest or any tender.

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

SHEWINGSFOUNDATION also reserves the right to negotiate or partly accept any or all the tenders received without assigning any reasons thereof. The tenderer may have to furnish rate Analysis for the scrutiny of rates by SHEWINGSFOUNDATION for negotiation etc .,if required. Tenders, which do not fulfill all or any of the mentioned conditions in the tender documents or are incomplete in any respect are liable to be rejected. Any discrepancies, omissions, ambiguities in the tender documents or any doubt as to their meaning should be reported in writing to The Chief Planning Officer, SHEWINGS FOUNDATION, UNITNO.C-403 04TH FLOOR TOWER C

TECHNOPARK PLOT NO 5, SECTOR 127, NOIDA, who will review the questions and where information sought is not clearly indicated or specified. SHEWINGSFOUNDATION will issue clarifications to all the tenderers, which will become part of the contract document. SHEWINGSFOUNDATION will not be responsible if the discrepancies, omissions, ambiguities in the Tender documents or any doubts as to their meaning are not brought to the notice of SHEWINGSFOUNDATION, before three working days prior to the date of submission of the tender.

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

For Supply and Installation of

DIGITAL MAMMOGRAPHY SYSTEM

Ref. No.: SWF/MAMMO/2026 Dated: 23TH SEPTEMBER 2026

Sl. No.	Particulars	Details
1	**Cost of Tender Documents**	Rs. 2,000/- (Demand Draft/ NEFT/RTGS in favour of SHEWingsFoundation) No fee if downloaded from official notice board. Account Details:- (NEFT /RTGS) Account No. 628405018293 ICICI Bank IFSC Code - ICIC006284 A/C Holder Name:- SHEWINGS FOUNDATION Type of Account- Current Account
2	**Earnest Money Deposit (EMD)**	Rs. 1,00,000/- in the form of Demand Draft in favour of SHEWINGS FOUNDATION.
3	**Tender Document Availability**	Available at the Office: C-403, 04th Floor, Tower C, Technopark, Plot No-5, Sector 127, Noida - 201313 or as per official notice.
4	**Last Date & Time for Submission**	16th October 2026; 14:00
5	**Tender Opening Date (Technical Bid)**	16th October 2026; 16:00
6	**Tender Opening Date (Financial Bid)**	16th October 2026; 17:00
7	**Venue of Tender Opening**	C-403, 04th Floor, Tower C, Technopark, Plot No-5, Sector 127, Noida - 201313

Note: *Any changes or further notifications in respect of the above Tender Documents shall be communicated through official notice boards / written communication. Respective bidders are advised to check regularly.*

Authorised Signatory
Cancer Screening & OPD Center
Sector 43, Noida, Gautam Buddha Nagar, U.P. - 201301

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

SECTION 1 – DEFINITIONS

The following terms shall have the meanings ascribed to them in this Tender Document:

Term	Meaning
Purchaser	Cancer Screening & OPD Center, Sector 43, Noida, Gautam Buddha Nagar, U.P. - 201301
Tenderer / Bidder	Manufacturer or Authorised Dealer/Distributor submitting a valid Tender
Supplier	Successful Bidder awarded the Purchase Order
PO	Purchase Order issued by the Purchaser
EMD	Earnest Money Deposit
PBG / PSB	Performance Bank Guarantee / Performance Security Bond
DD	Demand Draft
eLORA-AERB	Electronic Licensing Online and Related Applications - Atomic Energy Regulatory Board, Govt. of India
CMC	Comprehensive Maintenance Contract
AMC	Annual Maintenance Contract (Labour only; Spares extra)
GST	Goods and Services Tax
DICOM	Digital Imaging and Communications in Medicine

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

SECTION 2 – SCOPE OF WORK

Cancer Screening & OPD Center, Noida invites sealed tenders under a Two-Bid System (Technical Bid and Financial Bid) from eligible Manufacturers / Authorised Dealers for the following:

Sl.	Description of Item	Quantity	HSN Code
1	**Supply and Installation of Digital Mammography System (Model: MAM-VENUS+ / Series: VENUS DRV+ or equivalent)**	01 No.	90221490

The scope includes supply, transportation to the work site (Cancer Screening & OPD Center, Sector 43, Noida, Gautam Buddha Nagar, U.P. – 201301), installation, testing, commissioning, user training, and handing over of the equipment in full working condition.

SECTION 3 – ELIGIBILITY CRITERIA

Bidders must fulfil all of the following conditions to be considered eligible:

- Valid Manufacturer's Authorization Certificate (MAC) or Dealership Certificate from the OEM, specific to this Tender.
- High Quality Standards / ISO Certificate (ISO 9001 or equivalent) for the quoted product.
- Valid Company / Firm Registration Certificate (Certificate of Incorporation / Partnership Deed / Proprietorship Declaration).
- The unit should be European CE certified.
- The company should be ISO-9001:2015 certified with manufacturing, design of Flat panel detectors & X-Ray Generators.
- Valid GST Registration Certificate.
- PAN Card of the Firm or the Proprietor / Partner in whose name the firm is registered.
- Minimum average annual turnover of Rs. 1 Crore (Rupees One Crore) for the last three consecutive financial years. Audited Balance Sheets / CA Certificate to be submitted as proof.
- Affidavit on Non-Judicial Stamp Paper (minimum Rs. 10/-) attested by a Notary Public confirming that no vigilance / CBI / arbitration cases are pending and the firm has not been blacklisted by any sGovernment / Public institution.
- Copies of all Annexures (I through IV) duly filled and complete in all respects.
- Compliance with eLORA-AERB, Govt. of India requirements for supply and installation of ionising radiation equipment.
- Submission of valid EMD as specified in Section 5.

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

SECTION 4 – BID SUBMISSION SYSTEM

The Bid shall be submitted in a sealed outer envelope super-scribed:

"Tender for Supply and Installation of Digital Mammography System"
Ref. No.: SWF/MAMMO/2026 Dated: 23TH SEPTEMBER 2026

The sealed outer envelope must contain two separate sealed inner envelopes clearly marked as:

- Envelope 1 – Technical Bid (all technical documents, eligibility proofs, EMD, Annexures I to IV)
- Envelope 2 – Financial Bid (price schedule / Annexure V only)

The Financial Bid of only those bidders who are technically qualified shall be opened. The date and time of Financial Bid opening shall be communicated separately to technically qualified bidders.

SECTION 5 – EARNEST MONEY DEPOSIT (EMD)

- Every Tenderer shall submit an EMD of Rs. 1,00,000/- (Rupees One Lakh Only) in the form of a Demand Draft / Bank Guarantee / Fixed Deposit Receipt drawn in favor of "Cancer Screening & OPD Center, Noida", payable at Noida.
- The EMD shall be valid for a minimum period of 90 days from the date of Tender Opening.
- Tenders received without valid EMD shall be summarily rejected.
- EMD of unsuccessful Tenderers shall be refunded without interest within 30 days of finalization of the Tender.
- EMD of the successful Tenderer shall be retained until the Performance Security Bond (PSB) is furnished, after which it shall be refunded without interest.
- EMD shall be forfeited if the Tenderer withdraws or modifies the bid after submission, or fails to execute the contract after the award.

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

SECTION 6– TECHNICAL SPECIFICATIONS (ANNEXURE – I)

The following specifications are minimum mandatory requirements. Bidders must meet or exceed each specification. Deviations must be clearly declared in Annexure II (Technical Bid Format).

Parameter	Minimum Specification Required
1. MACHINE DESCRIPTION	
Model	Digital Mammography
Type	Fully Field Digital Mammography with 50 KHz High Frequency Microcontroller-based Generator
Application	Screening & Diagnosis
2. X-RAY GENERATOR	
Power Output	5 KW
Generator Frequency	High Frequency 50 kHz X-ray Generator
KVp Range	20 to 40 KVp in steps of 1 KVp
mA Output	Large Filament - 150 mA; Small Filament - 30 mA
mAs Range	1 - 700 mAs in steps
3. X-RAY TUBE HEAD	
X-Ray Tube	Rotating Anode X-Ray Tube
Focal Spot	Small Focus - 0.1 mm; Large Focus - 0.3 mm
Target Material	Molybdenum
Target Angle	10° (Small) & 16° (Large)
Inherent Filtration	0.5 mm Be
Anode Heat Storage Capacity	300 KHU
Anode Rotation Speed	3000 RPM
4. FLAT PANEL DETECTOR	
Type	Fixed Wired (FW)
Receptor Type	Amorphous Silicon TFT/PD Matrix Panel (Indirect Conversion)
Conversion Screen	CsI (Tl)
Detector Size	24 cm x 30 cm
Pixel Matrix (Resolution)	3500 (h) x 2800 (v)
Pixel Pitch	85 µm²

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

Limiting Resolution	Approx. 6 lp/mm
A/D Conversion	14 bits
Data Output	Gigabit Ethernet
5. CONTROL SYSTEM	
Control Panel	Fully integrated GUI Software controlled system
Operating Modes	APR Mode (auto exposure: Fatty/Normal/Dense breast); AEC Mode (Dual shot, automatic density selection)
Digital Display	KV, mAs, Error/Interlock messages, Filter Selected, Focal Spot Selected on LCD
6. COMPRESSION SYSTEM	
Compression Force	Up to 20 Kg ±2 Kg; Manual locking at 30 Kg
Compression Paddle Movement	Up to 180 ±10 mm
Compression Paddles	Normal size 24 x 30 cm; Square Spot Paddle with Magnification Device
Foot Switch	1 set for compression Up & Down movement
7. GANTRY	
Vertical Movement (Motor)	700 - 1250 mm (±10 mm) with 550 mm (±5 mm) travel
Angular Movement (Manual)	+180° to -180° (±2°) with locking mechanism
Source to Image Distance (SID)	650 mm (±10 mm)
Display	Digital display of angulation degree on LCD
8. MAGNIFICATION DEVICE	
Type	Light Weight Carbon Fiber with Frame
Magnification Factor	1.8X
9. X-RAY COLLIMATOR & RADIATION PROTECTION	
Collimator	Light Beam Collimator for 24 cm x 30 cm field size with Auto shut-off timer
Face Shield	Provided for Radiation Protection
Protective Barrier	Lead Glass Screen for Operator Protection
10. ELECTRICAL SPECIFICATION	
Input Voltage Supply	230 VAC ±10%, Single Phase
Frequency	50/60 Hz
Line Resistance	< 0.4 Ω
Earth-Neutral Voltage	≤ 3 VAC
11. ENVIRONMENTAL CONDITIONS – OPERATING	
Temperature Range	10°C to 40°C

Humidity	10% to 90%
Atmospheric Pressure	70 KPA to 106 KPA
12. ENVIRONMENTAL CONDITIONS – STORAGE	
Temperature Range	0°C to 40°C
Humidity	Maximum 0% at 90%
Atmospheric Pressure	70 KPA to 106 KPA
13. DISPLAY MONITOR	
Type	LCD/TFT
Screen Size	24 inches
Image Matrix	2MP
14. STANDARD ACCESSORIES (INCLUDED)	
Lead Aprons	0.25 mm lead equivalence - 02 Nos.
Dry Chemistry Printer	Two Tray Dry Chemistry Printer - 01 No.

Software Features

The system must include integrated imaging software with the following capabilities as a minimum:

Standard / X-Ray Generator Control:

- Dedicated imaging software for image acquisition with complete control of all image capture functions and exposure factors within the examination room.
- Enhances entire workflow by delivering diagnostic images instantly.
- Electronic transfer of X-Ray images to remote workstations, image archives, and printers.
- KV & mA increase/decrease; APR & AEC Mode selection; Focal Spot Large/Small selection; Exposure ON Indication.

Pre-Processing:

- Separate login for Manager and Technician.
- DICOM Modality Worklist for patient details; Manual & Emergency patient entry.
- Exam-specific algorithm image processing for consistent image quality.

Post-Processing:

- Image preview time less than 10 seconds.
- Image mirror & flip, rotation in steps and 90° steps; Dynamic Zoom with PAN; Negative Image; True Image & Fit to screen.
- Brightness & Contrast adjustment (Window Width & Window Level); DICOM Tags/Information view.
- EXI value indication for Normal, over/under exposure; Accept/Reject comments; Log list for Rejected, Accepted and Printed images.

Connectivity & Storage:

- Storage of images on CD/DVD with inbuilt DICOM viewer software viewable on any PC.
- DICOM 3.0 ready - compatible with PACS, RIS/HIS, and Dry Laser Printer.
- Print support on normal and DICOM printers; user-defined print settings; selection of different film sizes.

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

SECTION 7 – GENERAL TERMS AND CONDITIONS

7.1 Price & Taxes

- Prices shall be quoted F.O.R. Destination (Cancer Screening & OPD Center, Sector 43, Noida) inclusive of Basic Price and GST @ 12% (as applicable under Govt. of India notification, HSN Code: 90221490).
- In case of any change in GST rates by the Govt. of India, the revised rate prevailing at the time of dispatch shall be applicable.
- Any other Government levies / taxes applicable at the time of delivery shall be charged extra with prior written intimation.
- TDS is not applicable on sales transactions. Provisions of Section 194C or 194J of the Income Tax Act, 1961 are not applicable on payments against sale of equipment.

7.2 Delivery

- Delivery shall be effected within 10-11 weeks from the date of receipt of confirmed Purchase Order AND receipt of eLORA-AERB Procurement Permission from the hospital / facility authorities.
- The Purchaser shall provide the eLORA-AERB procurement permission as per applicable guidelines. The LD clock shall start only from the date of receipt of procurement permission.
- The Supplier shall be responsible for safe delivery, unloading, and placement of equipment at the designated site.

7.3 Delayed Penalty / Liquidated Damages (LD)

In case of delay beyond the stipulated delivery period, Liquidated Damages shall be levied on the Supplier as follows:

Delay Period (from Delivery Due Date)	Penalty (% of Total PO Value)
Up to 7 days	0.75%
8th Day to 15th Day	1.00%
16th Day to 22nd Day	3.00%
23rd Day to 30th Day	5.00%
Above 30 Days	10.00%

7.4 Warranty

- The equipment shall provide standard warranty of 3 years from the date of commissioning.
- During warranty, any component requiring replacement shall be replaced with a component of similar or superior specification at no cost to the Purchaser.
- Warranty shall not apply to damage caused by: misuse or lack of proper maintenance; tampering by unauthorized persons; external causes (liquid damage, fire, rodent bites, physical impact); consequential damages; natural disasters.

- In the event a critical component (e.g., X-Ray Tube) is discontinued by the OEM due to technological changes and is unavailable worldwide, a replacement with the new technology component shall be provided at an additional cost mutually agreed in writing.

7.5 Comprehensive Maintenance Contract (CMC)

- CMC charges: To be quoted by vendor.
- During CMC, any component requiring replacement shall be replaced with a component of similar specification type.
- The CMC contract must be signed within 30 days of the expiry of the warranty period. Failure to do so shall render the equipment out of warranty and CMC coverage; the rates quoted in this Tender shall not be binding thereafter.
- GST changes shall be applied at the prevailing rate at the time of commencement of the service contract.

7.6 Availability of Spare Parts

- The Supplier guarantees availability of spare parts for 10 years from the date of installation, or 10 years and 3 months from the date of dispatch, whichever is earlier.

7.7 eLORA-AERB Statutory Requirement

- As per the notification from eLORA-AERB, Govt. of India, the Purchaser is required to obtain procurement permission prior to delivery and installation of the equipment. The Supplier shall not deliver or install the equipment without such permission being granted and communicated.
- LD charges (if any) shall be calculated from the date of receipt of the duly granted procurement permission.

7.8 Force Majeure

No liability shall be attracted to the Supplier for non-performance or delayed execution of the order as a result of Force Majeure events including, but not limited to, war, fire, flood, epidemic, quarantine restrictions, acts of God, strikes, or any other event beyond reasonable control. The Supplier shall notify the Purchaser in writing within 7 days of the occurrence of such event.

7.9 Governing Law & Jurisdiction

- This Tender and resulting contract shall be governed by the laws of India.
- Legal proceedings, if any, shall fall within the jurisdiction of courts situated in Gautam Buddha Nagar (Noida), Uttar Pradesh.

7.10 Ethics & Anti-Collusion

Any attempt by a Tenderer to obtain confidential information, enter into unlawful agreements with competitors, or influence the Contracting Authority during the process of examining, clarifying, evaluating, and comparing tenders shall render the tender submitted by that Tenderer liable for outright rejection and forfeiture of EMD.

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

7.11 Complaint Registration

Post-installation complaints and service requests may be lodged at the Supplier's service contact numbers (to be furnished by the Supplier at the time of contract execution).

7.12 Right to Accept / Reject

Cancer Screening & OPD Center, Noida reserves the right to accept or reject any or all tenders in whole or in part without assigning any reason thereof. Placing of an order shall be presumed to be acceptance of all above terms and conditions by the Supplier.

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

ANNEXURE – II: TECHNICAL BID FORMAT

(To be submitted in Envelope 1 – Technical Bid, duly signed and stamped on each page)

Sl.	Item Description	Make & Model Quoted	Country of Origin	Catalogue / Compliance (YES/NO)	Warranty Period
1	Digital Mammography System (MAM-VENUS+ / VENUS DRV+ or equivalent)				

Note: All information in the above format is mandatory. Bidders must also submit a point-by-point compliance statement against each technical specification listed in Annexure I.

Signature of Authorised Signatory: _____
Name: _____
Designation: _____
Company Seal & Date: _____

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

ANNEXURE – III: DECLARATION BY BIDDER

I / We, _____(Name of Bidder / Firm), hereby declare that:

1. All information furnished in this Tender is true and correct to the best of my/our knowledge.
2. The Firm has not been blacklisted by any Government / Public Institution.
3. No vigilance / CBI / arbitration cases are pending against the Firm or its Principals.
4. I / We accept all Terms and Conditions laid down in this Tender Document.
5. I / We agree that the EMD shall be forfeited if I/We withdraw or modify the Bid after submission.
6. I / We understand that the Purchaser reserves the right to verify all information and documents submitted, and any misrepresentation shall lead to immediate disqualification and forfeiture of EMD.

Signature of Authorised Signatory: _____
Name: _____
Designation: _____
Company Seal & Date: _____

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

ANNEXURE – IV: EMD DETAILS

Particulars	Details (to be filled by Bidder)
DD / BG / FDR Number	
Amount (Rs.)	Rs. 1,00,000/-
Drawn on Bank	
Branch Name & Address	
Date of Issue	
Valid upto	
Payable at	Noida

Signature of Authorised Signatory: _____
Name: _____
Designation: _____
Company Seal & Date: _____

SHEWings Foundation
Address:- Unit No. C-403 04th Floor
Tower C, Technopark Plot No-5, Sector
127, Noida, G.B.Nagar, U.P - 201313
Call us: 0120-4074096
E: info@shewings.com
Website: www.shewingsfoundation.com

ANNEXURE – V: FINANCIAL BID (PRICE SCHEDULE)

(To be submitted ONLY in Envelope 2 – Financial Bid, separate from the Technical Bid)

Sl.	Item Description	Qty.	Unit Basic Price (Rs.)	GST @12% (Rs.)	Total Price Incl. GST (Rs.)
1	Digital Mammography System (MAM-VENUS+ / VENUS DRV+ or equivalent) – Supply & Installation at Cancer Screening & OPD Center, Sector 43, Noida	01			
Grand Total (inclusive of all taxes):					

Amount in Words: Rupees _____

HSN Code: 90221490 | Prices are inclusive of all costs for supply, installation, testing, commissioning, training, and F.O.R. Destination (Sector 43, Noida) delivery.

- CMC Rate (per year, after warranty): ____% of basic value, with ____% escalation each year.

Signature of Authorised Signatory: _____
Company Seal & Date: _____